

Calibration Certificate

Client Associated Steel Ind.
Ville Ste-Catherine, PQ
Model Number RC4069-2
Console SN# 11194
Software Version 4.0.5132
System Type Vehicle (Truck Scale)
Date Tested Oct 08, 2021
Certificate Expires Oct 08, 2022

THIS DOCUMENT #00007711
CERTIFIES THAT THIS
EQUIPMENT HAS BEEN TESTED
AND VERIFIED TO BE
FUNCTIONING AS PER RADCOMM
SYSTEMS PROCEDURES IN
CONFORMANCE WITH ISO
9001:2015 REQUIREMENTS. NIST
TRACEABLE SOURCES USED.



Calibrated and Tested

Parameters	Detector 1		Detector 2	
Detector SN	Detector n/a		Detector n/a	
Calibration Source used	2 uCi Cs137		2 uCi Cs137	
Background cps	2877		2853	
Other Test Source ?	9294		8924	
System Test cps	24967		24294	
Photomultiplier tubes	PMT1	PMT2	PMT1	PMT2
High Voltage Setting	980	0	720	0
Compton Edge Position (38 ± 1ch)	N/A	N/A	N/A	N/A
QC Snr Ratio (<2.2)	N/A	N/A	N/A	N/A
Visual Alarm	✓		✓	
Audio Alarm	✓		✓	
Vehicle Presence Sensors	✓		✓	

This certificate #00007711 certifies that the above equipment has been verified to function as per RadComm specifications on Oct 08, 2021
Serviced & Calibrated By Fred Walker



RADCOMM SYSTEMS CORP.
2931 Portland Drive
Oakville, Ontario, L6H 5S4
Tel: 905.829.8290
Fax: 905.829.1406
www.radcommsystems.com

Calibration Certificate

Client Associated Steel Ind.
Ville Ste-Catherine, PQ

Model Number RC4069-2

Console SN# 0164REC3

Software Version 4.0.5132

System Type Vehicle - Rail Wagon
Scale

Date Tested Oct 08, 2021

Certificate Expires Oct 08, 2022

THIS DOCUMENT #00007712
CERTIFIES THAT THIS
EQUIPMENT HAS BEEN TESTED
AND VERIFIED TO BE
FUNCTIONING AS PER RADCOMM
SYSTEMS PROCEDURES IN
CONFORMANCE WITH ISO
9001:2015 REQUIREMENTS. NIST
TRACEABLE SOURCES USED.



Calibrated and Tested

Parameters	Detector 1		Detector 2	
Detector SN	Detector 23335		Detector 23336	
Calibration Source used	2 uCi Cs137		2 uCi Cs137	
Background cps	2735		2537	
Other Test Source ?	8755		8292	
System Test cps	22952		22352	
Photomultiplier tubes	PMT1	PMT2	PMT1	PMT2
High Voltage Setting	800	0	780	0
Compton Edge Position (38 ± 1ch)	N/A	N/A	N/A	N/A
QC Snr Ratio (<2.2)	N/A	N/A	N/A	N/A
Visual Alarm	✓		✓	
Audio Alarm	✓		✓	
Vehicle Presence Sensors	✓		✓	

This certificate #00007712 certifies that the above equipment has been verified to function as per RadComm specifications on Oct 08, 2021

Serviced & Calibrated By Fred Walker



RADCOMM SYSTEMS CORP.
2931 Portland Drive
Oakville, Ontario, L6H 5S4
Tel: 905.829.8290
Fax: 905.829.1406
www.radcommsystems.com